

## Address for correspondence

Dr. C. W. Weykamp  
 Queen Beatrix Hospital  
 Beatrix Park 1  
 7101 BN Winterswijk  
 The Netherlands  
 Telephone: + 31 543 54 47 74  
 Fax: + 31 543 52 42 65  
 e-mail: c.w.weykamp@skbwinterswijk.nl  
 Chamber of Commerce: 09101380

Winterswijk, 2 September 2019

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
RV 2019-3, HbA1c K3A	<b>28,6</b>	1,0	<b>4,77</b>	0,09
RV 2019-3, HbA1c K3B	<b>32,0</b>	1,1	<b>5,08</b>	0,10

Samples arrived on 28 August and were analysed the next day.

Yours sincerely,



Dr. Cas Weykamp  
 Director



Carla Siebelder BSc.  
 Person in charge



Dr. Ir. R. J. Slingerland  
 chair

Dr. C. W. Weykamp  
 coordinator

Dr. H. van der Vuurst  
 treasurer